



204903-100

2018 年 9 月作成

## (RE) ノルディアN HbA1c 用 HbA1c コントロール

### 使用説明書

本品は、(RE) ノルディアN HbA1cを使用し、全血中のヘモグロビンA1c(HbA1c)を測定する際に、データ管理用として使用するものです。

#### 【使用法】

1. 本品1バイアルに精製水を正確に 1mL 加えて、約 10 分間ウェーブローターで混和するか、又は約 10 分間室温で静置後、緩やかに転倒混和してください。調製後の安定性は、冷蔵で 10 日間です(凍結不可)。なお、冷蔵保存した場合は、室内温度(15~30℃)に戻してから測定してください。
2. 2 時間以上静置した場合は、良く攪拌してから使用してください。
3. HbA1c(%)は、裏面に記載してあります。

#### 【取扱い上の注意事項】

1. 本品は、(RE) ノルディア N HbA1c 用のコントロールですので、本目的以外には使用しないでください。
2. 本品は、冷水で溶けにくい場合がありますので、20℃前後の精製水で溶解してください。
3. 本品は、HBs 抗原、HCV 抗体、HIV 抗体陰性を確認したヒト血液成分から調製しておりますが、使用の際には手袋などを着用し、検体と同様に感染の危険のあるものとして十分注意して取り扱ってください。

## HbA1c Control for (RE) NORUDIA N HbA1c kit

This product is the control for the assay to determine levels of hemoglobin A1c in whole blood using (RE) NORUDIA N HbA1c on an automated analyzer.

### Preparation

1. Add 1mL of purified water to one vial of this product. Mix the contents of the vial for approximately 10 minutes by the wave-rotor or leave the vial to stand for 10 minutes at room temperature and then gently mix by swirling and turning upside down several times. The reconstituted control is stable for 10 days if refrigerated. Avoid freezing. Allow the refrigerated control to reach room temperature (15-30°C) before measurement.
2. The reconstituted control kept for more than 2 hours should be mixed well before use.
3. HbA1c values are printed on the reverse.

### Handling Precautions

1. This product is the control specified for measurement with (RE) NORUDIA N HbA1c. Do not use this product for any other purpose.
2. Reconstitute this control with purified water at about 20°C as it may not be easily dissolved with cold water.
3. Human source material used in formulating this control has been confirmed to be negative for HBs antigens, HCV antibodies, HIV antibodies. However, handle this product in accordance with appropriate biohazard procedures as though capable of transmitting infectious agents. Wear disposable gloves to prevent disease transmission.

【包装 (Packaging)】 1mL 用×2 濃度×6 (Two concentrations x 6 vials for 1mL)

【貯蔵方法 (Storage Temperature)】 2~10℃ (2-10 °C)

【使用期限 (Expiration Date)】 外装に表示してあります。(The expiration date is printed on the outer packaging.)

#### 製造元

積水メディカル株式会社  
東京都中央区日本橋二丁目1番3号

#### Manufacturer

SEKISUI MEDICAL CO.,LTD.  
1-3, Nihonbashi 2-chome, Chuo-ku, Tokyo 103-0027 JAPAN

(RE) ノルディアN HbA<sub>1c</sub>用 **HbA<sub>1c</sub> コントロール 濃度表**

**Table for the concentration of HbA<sub>1c</sub> Control** for (RE) NORUDIA N HbA<sub>1c</sub> kit

**Lot No. 122RIU**

HbA<sub>1c</sub>(%)はキットのロットごとに異なります。ご使用の際には、必ずバイアル瓶ラベルに記載されている製造番号と上記記載の製造番号が一致することをご確認ください。

The HbA<sub>1c</sub> values are different from each production lot. Please confirm that the labeled Lot No. of each vial is consistent with the Lot No. in this sheet before testing.

	HbA <sub>1c</sub>	
	NGSP 値(%)	JDS 値(%)
コントロール Low HbA <sub>1c</sub> Control Low	4. 9—5. 5	4. 6—5. 1
コントロール High HbA <sub>1c</sub> Control High	9. 4—10. 4	9. 0—9. 9